

CERTIFICATE

Survey of 09 September 2021

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Haemostasis 18 - POCT (286):

Validity 12 months:

ACT

(R) analysis is subject to the RiliBÄK

Participant:
8195
MUDr. Marie Kloudova
INLAB Medical, s.r.o.
Kostelec u krizku275
25168 Kamenice

Düsseldorf, 20 September 2021



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Andreas Giebl
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 09 September 2021

You have participated in the External Quality Assessment with the following analysis

Haemostasis 18 - POCT (286):

ACT

(R) analysis is subject to the RiliBÄK

Participant:
8195
MUDr. Marie Kloudova
INLAB Medical, s.r.o.
Kostelec u krizku275
25168 Kamenice

Düsseldorf, 20 September 2021



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Andreas Giebl
(Adviser)

Listing and evaluation of the results

8195: MUDr. Marie Kloudova
INLAB Medical, s.r.o.


Survey of 09 September 2021

Adviser: **Andreas Giebl**
Klinikum Augsburg
Institut für Transfusionsmedizin und
Hämostaseologie
Stenglinstraße 2
86156 Augsburg

Tel.: +49 821 400 2668
Mail: andreas.giebl@klinikum-augsburg.de

286

Haemostasis 18 - POCT

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
ACT	(1) 51	seconds	101	124	SV	80.8	167	-18.5%	-1,06	+	
	(1) 52		180	204	SV	146	262	-11.8%	-0,83	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device	Testkit	Time of measureme nt	recalcificatio n
ACT	51		101	1.00		AB	AB01	199		
	52		180	1.00						

(1) In individual cases is a statistically valid valuation with consensus value not given, because size of the collective < 8 values.

Individual summary of results

8195: MUDr. Marie Kloudova
INLAB Medical, s.r.o.

Survey of 09 September 2021

286

Haemostasis 18 - POCT

ACT (seconds, N = 76)										
Collective	Sample	Target value	Target range		Participants collective			Rate (%)		
					AVG	CV	Num.	Sam.	total	
Keller, ACT low range	51	178	153	-	203	178	6.97	20	100	100
	52	213	186	-	240	213	6.34	20	100	100
TK.195, 197	51	112	78.6	-	155	112	14.9	23	87.0	87.0
	52	168	116	-	220	168	15.6	23	95.7	95.7
Medtronic, HR Kartusche	51	122	104	-	140	122	7.39	12	91.7	91.7
	52	171	130	-	212	171	11.9	12	91.7	91.7
TK.199, 232-233 (1)	51	124	80.8	-	167	124	17.4	4		
	52	204	146	-	262	204	14.2	4		
Abbott, ACT Kaolin	51	117	98.0	-	129	117	5.13	17	94.1	94.1
	52	189	153	-	225	189	9.52	17	94.1	94.1

(1) In individual cases is a statistically valid valuation with consensus value not given, because size of the collective < 8 values.

Rate of success: 93,4%

